



# OFFICE OF THE CHIEF DISTRICT MEDICAL & PHO- cum-DMD, CUTTACK



## Corrigendum Notice

Notice No: - 9176 / Tender.

Dated: - 20 / 07 / 2020

With reference to the letter no 9064 dated 16.07.2020 and the tender call notice no 9063/CDM&PHO/2020-21 dated 16.07.2020 the technical specification of pulse oxymeter is hereby modified and new specification is uploaded. Potential bidders are requested to submit the tender as per the new specification of pulse oxymeter.

<u>Sl No</u>	<u>Name of the items(EIF)</u>	<u>Specification</u>	<u>Requirement</u>
1	Finger tip pulse Oxymeter	<p>TECHNICAL SPECIFICATIONS</p> <ol style="list-style-type: none"><li>1. The Pulse oximeter, based on all digital technology is intended for noninvasive spot-check measurement of functional oxygen saturation of arterial hemoglobin (SpO2). Advanced DSP algorithm can reduce the influence of motion artefact and improve measurement accuracy of low perfusion.</li><li>2. 2 Color OLED display.</li><li>3. Display SpO2, PR, PI, SpO2 waveform (pleth or line).</li><li>4. Display signal strength.</li><li>5. Four display modes convenient to overview the data.</li><li>6. Low voltage indicator, Low power consumption.</li><li>7. 2 AAA alkaline batteries.</li><li>8. Light, compact less than 50 grams.</li><li>9. Product should be ISO &amp; CE certified, Made in India product will be given competitively.</li><li>10. 1 year replacement warranty</li></ol>	2000Nos

*SR*  
20.7.20

Chief District Medical &  
PHO- Cum-DMD, NHM, Cuttack